

# Certificate of Analysis List

For request number 594195

<u>Catalog Number Entered</u>	<u>Lot Number Entered</u>	<u>Related Catalog Number</u>	<u>Related Lot Code</u>	<u>Description</u>
LZW9711.99 uS/cm	4057	N/A	N/A	Conductivity Standard Solution, 1413

---

Total Enclosures: 1

**APPROVED**

*By Antonio Tavares at 4:05 pm, Apr 23, 2014*



An ISO 9001 Certified Company

*Certificate of Analysis*

COMMODITY: **Conductivity Standard Solution, 1413 uS/cm**

COMMODITY NUMBER: **LZW9711.-99** MANUFACTURE DATE:

DATE OF ANALYSIS:

LOT NUMBER: **A4057**

**2/24/2014**

**2/24/2014**

---

<i>TEST</i>	<i>SPECIFICATIONS</i>	<i>RESULTS</i>
Conductivity	1398 to 1426 uS/cm	1412.0 uS/cm

The expiration date is Feb 2015

Certified by \_\_\_\_\_

A handwritten signature in cursive script that reads "Scott Als".

Scott Als  
Analytical Services Chemist